

IAP20 Rec'd PGT/PTO 14 DEC 2005

Method for Fluorescence Microscopy

The invention pertains to a method in fluorescence microscopy, especially laser scanning microscopy.

Prior Art

A conventional application of light microscopy for investigation of biological preparations is fluorescence microscopy (lit.:

Pawley, "Handbook of biological confocal microscopy", Plenum

Press, 1995). Specific dyes are used here for specific marking of tissues, cells Method Using Variable Dyes and/or cell parts

or also materials. Dye Combinations for Image Evaluation

Cross-Reference to Related Applications

The present patent application is a nationalization of International Application No. PCT/EP2004/006202, filed June 9, 2004, which is based on, and claims priority from, German Application No. DE 103 27 382, filed June 16, 2003, both of which are incorporated herein by reference in their entireties.

Background of the Invention

Field of the Invention

The invention relates to a method used in fluorescence microscopy, especially laser scanning microscopy, in general, and to the use of dyes for image evaluation, in particular.

Description of Related Art

A conventional application of light microscopy for investigation of biological preparations is fluorescence microscopy which is discussed in the "Handbook of Biological Confocal Microscopy, Second Edition", Pawley, Plenum Press, 1995. Specific dyes are used in fluorescence microscopy for specific marking of tissues, cells, cell parts, or other materials.

The emitted photons of a specific energy excite the dye molecules from the ground state to an excited state by absorption of a photon. This excitation is generally referred to as one-photon absorption (See Figure 1a). The dye molecules so excited can return in different ways to the ground state. In fluorescence microscopy the transition with emission of a fluorescence photon is the most important. The wavelength of the emitted photon is generally red-shifted based on the Stokes shiftShift in comparison to the excitation radiation and therefore has a greater wavelength. The Stokes shiftShift permits separation of the fluorescence radiation from the excitation radiation.

A multiphoton excitation is shown in Figure 1b.

__The fluorescence light is split and observed separately from excitation radiation with appropriate dichroic beam the

splitters in combination with block filters. The depiction of individual cell parts stained with different dyes is possible on this account. However, several parts of a preparation, in principle, can also be simultaneously stained with different, dyes by specifically adding dyes <u>fresulting in multiple</u> fluorescence. Special dichroic beam splitters are again used to distinguish the fluorescence signals emitted by the individual dyes.

The prior art is to be explained belows follows based on the example of a confocal laser scanning microscope (LSM) (such as that show schematically in Figure 2).

___An LSM is divided into essentially four modules: light source L, scan module_S, detection unit D_and microscope_M. These modules are further described below. DE 19702753 Al is additionally referred to. U.S. Patent No. 6,167,173 also provides a detailed explanation of the LSM and is incorporated by reference herein as if reproduced in its entirety.

____For specific excitation of different dyes in one preparation, different wavelengths are used in an LSM laser. The choice of excitation wavelength is guided according to the absorption properties of the dyes being investigated. The excitation radiation is generated in the light source module_L. Different lasers 13.1 and 13.2 are used here (for example,

argon, argon-krypton, TiSa lasers). Selection of the wavelength and adjustment of the intensity of the required excitation wavelength additionally occurs in the light source module L, for example by using an acousto-optic crystal 15.1 and 15.2. The laser radiation then goes to the scan module S via a fiber 14.1 and 14.2 and/or appropriate mirror arrangement.

_____The laser radiation generated in the light source is focused into the preparation with limited diffraction with objective (2)—4 via the scanner 23, scan optics 16 and 17 and tube linelens 9. The focus scans the sample in the x-y direction point-like. The pixel residence times during scanning over the sample are generally in the range of less than a microsecond to a few seconds.

During confocal detection (descanned detection) of the fluorescence light, the light emitted from the focal plane (specimen) 5 and from the overlying and underyingunderlying planes goes to a dichroic beam splitter (MDB) 24 via the scanner 23. This separates the fluorescence light from the excitation light. The fluorescence light is then focused on a diaphragm (confocal diaphragm/pinhole+) 29, which is situated precisely in a plane conjugated to the focal plane. Fluorescence light fractions outside of the focus are suppressed on this account.

By varying the diaphragm size the optical resolution of the microscope can be adjusted. Behind the diaphragm there is an additional dichroic block filter (EF) 30 that again suppresses the excitation radiation. After passing through the block filter, the fluorescence light is measured by means of a point detector (PMT+) 31.

During use of multiphoton absorption, excitation of dye fluorescence occurs in a small volume on which the excitation intensity is particularly high. This region is only slightly larger than the detected region during use of a confocal arrangement. The use of a confocal diaphragm can therefore be omitted and detection can occur directly after the objective +5 in the nondescanned detection region 71 of the microscope.

____In another arrangement for detection of dye fluorescence excited by multiphoton absorption, descanned detection again occurs, but this time the pupil of the objective is imaged in the detection unit (nonconfocal descanned detection+) D.

Only the plane (optical section) that is situated in the focal plane of the objective is detected by a three-dimensionally illuminated image by both detection arrangements in conjunction with the corresponding one photon or multiphoton absorption. By marking several optical sections in the x-y plane

at different depths z of the sample, a three-dimensional image of the sample can then be generated in computer-controlled fashion. The LMS is therefore suitable for investigating thick preparations. The excitation wavelengths are determined by the employed dye with specific absorption properties. Dichroic filters 30 adjusted to the emission properties of the dye ensure that only the fluorescence light emitted by the corresponding dye is measured by the point detector.

In biomedical applications several different cells or cell regions are now marked with different dyes simultaneously (multifluorescence). The individual dyes can be detected separately with the prior art based either on different absorption properties or emission properties (spectra). For this purpose additional splitting of the fluorescence light from several dyes occurs with the secondary beam splitters (DBS) 28 and a separate detection of the individual dye emissions in separate point detectors (PMT $x \rightarrow 21$) 31. Flexible adjustment of detection and excitation to corresponding new properties by the user is not possible with the arrangement described above. New dichroic beam splitters and block filters must instead be created for each (new) dye.

_____If the emission spectra of two dyes overlap, the previous detection devices reach their limits. In order to avoid overlap between two dyes, the spectral detection range must be restricted. The range in which the two dyes overlap is simply cut out for this purpose and not detected. The efficiency of the detection unit therefore deteriorates. An equal signal-to-noise ratio can only be achieved by increasing the excitation power, through which preparation damage can occur. Nowadays a maximum of up to six different dye probes are therefore simultaneously used, since the dyes otherwise could not be separated owing to the strongly overlapping emission bands.

Previously dyes have been modified so that they either differ from each in their absorption properties or in their emission properties. Figure 3 shows the emission spectra of different typical dyes. The emission signal is plotted as a function of wavelength. The dyes denoted 1 to 4 differ in position and form of their emission spectra. These dyes, are in most cases toxic for living preparations. Investigations on Investigation of the evolution of structure in living preparations are is therefore impossible. In the 90s, dyes occurring in nature, so-called fluorescing proteins (GFP, YFP, CFP, TOPAS, GFT, RFP) discovered (company: by Clonetech, USA).

Mountain View, CA www.clontech.com. Fluorescence dyes for specific marking of preparations are used in all of the aforementioned systems.

P70986US0

In the Zeiss laser scanning microscope META the fluorescence is spit spectrally. For this purpose the emission layer is split from the excitation light in the scan module or in the microscope (during multiphoton absorption) by means of the main color divider (MDB). A block diagram of the following detector unit is shown in Figure 5. The light of the sample is focused by means of an imaging optics PO during confocal detection through an iris (pinhole) PH so that fluorescence that originated outside of the focus is suppressed. During nondescanned detection the diaphragm drops out. The light is now broken down into its spectral fractions by means of an angle-dispersive element DI-Brief Summary of the Invention

The present invention covers a method using fluorescence microscopy for image evaluation using a laser scanning microscope in which an at least partially spectrally resolved detection of the fluorescence spectrum occurs. Reference spectra are used for spectral demixing. Temporally and/or spectrally variable dyes and/or dye combinations are employed for recording of the reference spectra. Finally, the recorded reference spectra are inspected for image evaluation.

The inventive method records organic, intracellular, and intercellular processes. Cells and/or cell populations can also

be recorded. Reference spectra of photoconvertible dyes, photoactivatable dyes, and indicator dyes can also be recorded.

Of interest to the inventive method is that dyes used in recording reference spectra change their spectra dynamically based on intracellular processes. Also, dyes that have a different rise in fluorescence intensity are used. One fluorescing protein that has been used to practice the present invention is Kaede.

Brief Description of the Several Views of the Drawings

Figure 1a schematically illustrates the concept of one-photon absorption.

Figure 1b schematically illustrates the concept of multiphoton absorption.

Figure 2 is a schematic drawing of a laser scan microscope used in practicing the inventive method.

Figure 3 is a chart showing the emission spectra of several typical dyes.

Figure 4 schematically depicts how different image channels are formed.

11 0 200 1/11 15 / 0/

Figure 5 is a block diagram of the main color divider (MDB) used in practicing the inventive method.

Detailed Description of the Invention

In describing preferred embodiments of the present invention illustrated in the drawings, specific terminology is employed for the sake of clarity. However, the invention is not intended to be limited to the specific terminology so selected, and it is to be understood that each specific element includes all technical equivalents that operate in a similar manner to accomplish a similar purpose.

In a laser scanning microscope, such as the Zeiss META, the fluorescence is spit spectrally. For this purpose the emission layer is split from the excitation light in the scan module S of in the microscope (during multiphoton absorption) by means of the main color divider (MDB) 24. A block diagram of a detector unit used in practicing the present invention is shown in Figure 5.

The light of the sample is focused by means of an imaging optics 50 during confocal detection through an iris (pinhole) 52 so that fluorescence that originated outside of the focus is suppressed. During nondescanned detection the diaphragm drops out. The light is now broken down into its spectral fractions by

P70986US0

means of an angle-dispersive element. Prisms, gratings and acousto-optic elements are considered as angle-dispersive elements. The light split by the dispersive element into its spectral components is finally imaged in a line detector DE.56. This line detector DE56 therefore measures the emission signal as a function of wavelength and converts it to electrical signals. In addition, the detection unit can have a line filter (not shown) connected in front to suppress excitation wavelengths.

The depicted structure essentially describes a Cerny—Turner—design.—Design. During confocal detection the light ±60 of the sample is focused with the pinhole optics ±050 through the confocal diaphragm ±1.52. During a nondescanned detection, in the case of multiphoton absorption, this diaphragm can drop out. The first imaging mirror \$162 collimates the fluorescence light. The light then impinges on a line grating 664, for example a grating with a line number of 651 lines per mm. The grating diffracts the light according to its wavelength in different directions. The second imaging mirror \$266 focuses the individual spectrally split wavelength fractions onto the corresponding channels of the line detector ±56. The use of a line-secondary electron multiplier of H7260 made by the

Hamamatsu Co. H7260—Photonics, K.K., headquartered in Hamamatsu City, Japan (www.hamamatsu.com), is particularly advantageous. The detector has 32 channels and high sensitivity. The free spectral range of the aforementioned variant is about 350 nm. The free spectral range in this arrangement is uniformly distributed in the 32 channels of the line detector so that an optical resolution of about 10 nm is produced. This arrangement is only conditionally suitable for spectroscopy. However, its use in an imaging system is advantageous, since the signal is still relatively large per detection channel because of the relatively broad detected spectral band. A shift of the free spectral region can additionally occur by rotation of the grating.

Another possible variant <u>could_involve_involves</u> the use of a matrix detector (for example CCD). In this case splitting into different wavelength fractions is carried out in one coordinate by the dispersive element. In the remaining direction, a complete line (or column) of the scanned image is imaged on the matrix detector. This variant is particularly advantageous in constructing a line scanner (lit.: Corle, Kino: "Confocal Scanning Optical Microscopy and Related Imaging Systems," Academic Press, 1996). The essential design corresponds

essentially to that of an LSM according to Figure 2. However, instead of a point focus, a line is imaged in the focus and the sample being investigated is only scanned in one direction. A slit diaphragm instead of a perforated diaphragm serves as confocal diaphragm in such a structure. A nondescanned detection during use of multiphoton absorption can also occur with this arrangement. The confocal diaphragm can again drop out for this purpose.

By spectral splitting of the fluorescence light, after recording of the fluorescence spectra of the fluorescence markers in pure form and recording of the spectra with fluorescence fractions of several markers, a separate recording of the spectral fractions can occur by an unmixing method (such as that disclosed in DE 19915137 Al).

Invention:

It has now been recognized according to the invention that the number of fluorescence markers used for fluorescence marking can be reduced or combinatorics can be utilized, if not only fluorescence spectra are to be used in pure form as reference spectra but also reference spectra of mixed forms are recorded. These mixed forms can be characterized, for example, by the time-dependent color state of the biological material, if a fluorescence marker slowly leads to discoloration.

In addition, such mixed states can be characterized by a mixed

color if a fluorescence marker changes it color or its excitation properties.

WO 2004/1:13987 · · · 14/5	PCT/EP2004/006202
The state of the s	P70986US0
Such mixing conditions can be	e produced in different ways:
they can be present in the sample,	be produced by irradiation of
the sample, or be the result of a	biological process excited by
irradiation.	
Mixed spectra can characteri	ze a biological process, for
example, a concentration change,	in which a first spectrum
corresponds to a lower concentra	tion state and at least one
additional spectrum corresponds to	a higher concentration state.
By means of the different reference are defined and correspondingly	
such references can occur or	ver the entire image or
advantageously over marked "regi	ions of interest" (ROI). A
deliberate manipulation by define	d irradiation can also occur
over such ROI. A reference can a	lso be determined in a first
region and deliberate irradiation	and measurement can occur in
an additional region by extractio	on of mixed spectra. Demixing
and depiction of the spectra can	occur after imaging recording
or during imaging recording.	
It was surprisingly found	that new markers like the
fluorescing protein Kaede, which t	urns from green to red during

irradiation (lit.: Ando, R., Harna, H., Yamamoto-Hino, M.,

Mizuno, H. and Miyawaki, A. (2002), An optical marker based on

the UV-induced green-to-red photoconversion of a fluorescent protein, PNAS 99/20, 12651-12656) during recording of organic processes, for example inter- and intracellular processes can be used for marking of individual cells or cell populations and spectrally detected. Photoconvertible dyes that change their spectra dynamically based on intracellular processes or dyes that are used for FRET, but also other indicator dyes can be advantageously used by the method according to the invention.

____For example, different cells or cell groups in a cell population that has been marked with the dye Kaede can be exposed and converted for different lengths of time with UV or violet light. Different color mixing ratios are then established that are recorded as reference spectra. ___Different cell populations can then be individually recorded over time. This is true not only for cell cultures, but can also concern both subcellular structures and also entire organisms, in which a certain cell population is irradiated and can be observed in its development, in which the resulting reference can be assigned to specific image channels and can therefore be followed.

A corresponding region can be selected over the ROI. An analysis of transport processes on the cellular and subcellular level can occur. Advantageously, still only one dye is used,

which is placed in different states via radiation of other effects that are clearly identifiable by reference formation.

A different rise in fluorescence intensity, as during PA-(photoactivatable GFP, lit.: Patterson, G. H. GFP and Lippincott-Schwartz, J. (2002), A photoactivatable GFP for selective photobleaching of proteins and cells, Science 297, 1873-1877) after excitation with violet light can serve as reference. In the Zeiss LSM META ROIs can be interactively defined directly in the image. The selected laser is switched on and off with pixel precision at the boundary of these regions.

In time series dialog, the start and end of the time series as well as intervals and delays between recordings are fixed. The irradiation parameters that lead to a change in dye properties by photoactivation or photoconversion, for example, repetition rate, wavelength, intensity, position can automatically incorporated. Evaluation can occur after the experiment or on-line during recording in order to be able to intervene directly in the course of the experiment. The average intensities of ROI in time, as well as the times photoactivations and conversions are indicated.

The META detector permits the entire spectrum of emission, for example of Kaede, to be recorded and to separate the

WO 2004/113987	17/5	PCT/EP2004/006202
		P70986US0
corresponding mixed	forms spectrally during	measurement and to
display the demixed	channels.	
Figure 4 schema	tically depicts how diffe	erent image channels
CH1-CH3 are formed,	in which, as shown, diffe	erent spectral mixed
distribution of CH1-	3 are used as reference	and referred to for
image evaluation.		

5 65 g